

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### **Rabbit anti Chicken IgG (heavy and light chains), Clone: [Polyclonal], Monoclonal NMB-RACH/IGG(H+L)**

|                          |  |
|--------------------------|--|
| Artikelname              | Rabbit anti Chicken IgG (heavy and light chains), Clone: [Polyclonal], Monoclonal  |
| Artikelnummer            | NMB-RACH/IGG(H+L)  |
| Hersteller Artikelnummer | RACH/IgG(H+L)  |
| Alternativnummer         | NMB-RACH/IGG(H+L)  |
| Hersteller               | NordicMubio  |
| Wirt                     | Rabbit   |
| Kategorie                | Antikörper   |
| Produktbeschreibung      | The reactivity of the antiserum is directed to the Fc and Fab subunits of the IgG molecule. It includes a certain degree of reactivity with other immunoglobulins via the common Fab portion. It does not react with any non-Ig protein in chicken serum, ...  |
| Klonalität               | Monoclonal   |
| Klon-Bezeichnung         | [Polyclonal]   |
| Puffer                   | Delipidated, heat inactivated lyophilized stable whole serum. No preservative added, as it may interfere with the antibody activity. No foreign protein added. Total protein and IgG concentration in the antiserum are comparable to those of pooled rabbit s |
| Quelle                   | Purified normal IgG isolated from pooled chicken serum. Freund's complete adjuvant is used in the first step of the immunization procedure.  |

|                        |  |
|------------------------|--|
| Formel                 | Delipidated, heat inactivated lyophilized stable whole serum. No preservative added, as it may interfere with the antibody activity. No foreign protein added. Total protein and IgG concentration in the antiserum are comparable to those of pooled rabbit s |
| Anwendungsbeschreibung | Precipitation assays. In immunoelectrophoresis use 2 µl or equivalent against 120 µl antiserum. In double radial immunodiffusion (Ouchterlony) use a rosette arrangement with 10 µl antiserum in a 3 mm diameter centre well and 2 µl serum samples (neat and  |