

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### **AMPIPROBE NG/CT/TV Assay Kit ENZ-ENZ-GEN207-0100**

Artikelname	AMPIPROBE NG/CT/TV Assay Kit
Artikelnummer	ENZ-ENZ-GEN207-0100
Hersteller Artikelnummer	ENZ-GEN207-0100
Alternativnummer	ENZ-ENZ-GEN207-0100-100
Hersteller	Enzo Life Sciences
Kategorie	Kits/Assays
Spezies Reaktivität	Bacteria, Parasite
Produktbeschreibung	AMPIPROBE NG/CT/TV Assay Kit...
Sensitivitaet	In a validation study approved by the New York State Department of Health, the AMPIPROBE NG/CT/TV Assay was determined to have the following percentage of Sensitivity: Neisseria gonorrhoeae - 100.00% Chlamydia trachomatis - 100.00% Trichomonas vaginalis
Target-Kategorie	Neisseria gonorrhoeae  Chlamydia trachomatis  Trichomonas vaginalis
Anwendungsbeschreibung	The AMPIPROBE NG/CT/TV Assay is compatible with any properly calibrated qPCR thermal cycler capable to detect fluorescence decay. It has been validated for use on the QIAGEN Rotor-Gene Q