

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

AMPIPROBE Ureaplasma spp. / M. genitalium / M. hominis Assay Kit ENZ-ENZ-GEN209-0100

Artikelname	AMPIPROBE Ureaplasma spp. / M. genitalium / M. hominis Assay Kit
Artikelnummer	ENZ-ENZ-GEN209-0100
Hersteller Artikelnummer	ENZ-GEN209-0100
Alternativnummer	ENZ-ENZ-GEN209-0100-100
Hersteller	Enzo Life Sciences
Kategorie	Kits/Assays
Spezies Reaktivität	Bacteria
Produktbeschreibung	AMPIPROBE Ureaplasma spp. / M. genitalium / M. hominis Assay Kit...
Sensitivitaet	In a validation study approved by the New York State Department of Health, the AMPIPROBE Ureaplasma spp. / M. genitalium / M. hominis Assay was determined to have the following percentage of Sensitivity: Ureaplasma spp. - 97.80% M. genitalium - 87.30 M.
Target-Kategorie	Ureaplasma spp. Mycoplasma genitalium Mycoplasma hominis
Anwendungsbeschreibung	The AMPIPROBE Ureaplasma spp. / M. genitalium / M. hominis Assay is compatible with any properly calibrated qPCR thermal cycler capable to detect fluorescence decay. It has been validated for use on the QIAGEN Rotor-Gene Q