

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### Sheep anti Guinea Pig IgA (Fc specific), Clone: [Polyclonal], Monoclonal NMB-SHAGP/IGA(FC)

|                          |  |
|--------------------------|--|
| Artikelname              | Sheep anti Guinea Pig IgA (Fc specific), Clone: [Polyclonal], Monoclonal   |
| Artikelnummer            | NMB-SHAGP/IGA(FC)  |
| Hersteller Artikelnummer | ShAGp/IgA(Fc)  |
| Alternativnummer         | NMB-SHAGP/IGA(FC)  |
| Hersteller               | NordicMubio  |
| Wirt                     | Sheep  |
| Kategorie                | Antikörper   |
| Produktbeschreibung      | The reactivity of the antiserum is restricted to the Fc part of the IgA molecule. In immunoelectrophoresis and radial immunodiffusion, using various antiserum concentrations against normal guinea pig serum a single precipitin line is obtained which s...  |
| Klonalität               | Monoclonal   |
| Klon-Bezeichnung         | [Polyclonal]   |
| Puffer                   | Delipidated, heat inactivated, lyophilized, stable whole serum. No preservative added. Total protein and IgG concentrations in the antiserum are comparable to those of pooled normal sheep serum. No foreign proteins added. Reconstitute the lyophilized ant |
| Quelle                   | Highly purified normal IgA isolated from pooled guinea pig serum. Freund's complete adjuvant is used in the first step of the immunization procedure.  |

|                        |   |
|------------------------|---|
| Formel                 | Delipidated, heat inactivated, lyophilized, stable whole serum. No preservative added. Total protein and IgG concentrations in the antiserum are comparable to those of pooled normal sheep serum. No foreign proteins added.   |
| Anwendungsbeschreibung | Precipitation assays. In immunoelectrophoresis use 2 $\mu$ l serum or equivalent against 120 $\mu$ l antiserum. In double radial immunodiffusion (Ouchterlony) use a rosette arrangement with 10 $\mu$ l antiserum in 3 mm diameter center well and 2 $\mu$ l serum samples (neat |