

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### Chlorate Anion IC Standard from Sodium Chlorate at 100 µg/mL in H<sub>2</sub>O HPS-IC-CLO3-500

Article Name	Chlorate Anion IC Standard from Sodium Chlorate at 100 µg/mL in H <sub>2</sub> O
Biozol Catalog Number	HPS-IC-CLO3-500
Supplier Catalog Number	IC-CLO3-500
Alternative Catalog Number	HPS-IC-CLO3-500-500
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Chlorate 100 µg/mL in H <sub>2</sub> O Exp Date: 1 Year 18 months expiry date. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:2016 accredited....