

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Nitrate Anion IC Standard from Sodium Nitrate at 10,000 µg/mL in H₂O HPS-IC-NO-10M-500

Article Name	Nitrate Anion IC Standard from Sodium Nitrate at 10,000 µg/mL in H ₂ O
Biozol Catalog Number	HPS-IC-NO-10M-500
Supplier Catalog Number	IC-NO-10M-500
Alternative Catalog Number	HPS-IC-NO-10M-500-500
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Nitrate 10,000 µg/mL in H ₂ O 18 months expiry date. Traceable to NIST 3185. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:2016 accredited....