

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### **Arsenic (As) ICP Standard at 1,000 µg/mL in 2% HNO<sub>3</sub>** **HPS-10003-1-100**

Article Name	Arsenic (As) ICP Standard at 1,000 µg/mL in 2% HNO <sub>3</sub>
Biozol Catalog Number	HPS-10003-1-100
Supplier Catalog Number	10003-1-100
Alternative Catalog Number	HPS-10003-1-100-100
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	As 1000 µg/mL in 2% HNO <sub>3</sub> 18 months expiry date. Traceable to NIST SRM 3103. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:20166 accredited....