

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Copper (Cu) ICP Standard at 1,000 µg/mL in 2% HCl **HPS-100014-2-100**

Article Name	Copper (Cu) ICP Standard at 1,000 µg/mL in 2% HCl
Biozol Catalog Number	HPS-100014-2-100
Supplier Catalog Number	100014-2-100
Alternative Catalog Number	HPS-100014-2-100-100
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Cu 1000 µg/mL in 2% HCl, 18 months expiry date. Traceable to NIST SRM 3114. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:20166 accredited....