

Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Dangerous goods 16-component Contract Laboratory Program (CLP) ICP Calibration Standard 1 Two solutions (A & B) Each solution Contains: HPS-CLP-CAL-1-100

Article Name	Dangerous goods 16-component Contract Laboratory Program (CLP) ICP Calibration Standard 1 Two solutions (A & B) Each solution Contains:
Biozol Catalog Number	HPS-CLP-CAL-1-100
Supplier Catalog Number	CLP-CAL-1-100
Alternative Catalog Number	HPS-CLP-CAL-1-100-100
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Calibration Std 1 in 4% HNO3 16 comp, (μg/mL) Solution A: Beryllium at 50, Chromium at 200, Copper at 250, Cobalt, Manganese, Nickel, Vanadium, Zinc at 500, Iron at 1,000, Aluminium, Barium at 2,000, Calcium, Potassium, Magnesium, Sodium at 5,000 4%