

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### **Nirite Anion IC Standard from Sodium Nitrate at 1,000 µg/mL in H2O HPS-IC-N-M-100**

Article Name	Nirite Anion IC Standard from Sodium Nitrate at 1,000 µg/mL in H2O
Biozol Catalog Number	HPS-IC-N-M-100
Supplier Catalog Number	IC-N-M-100
Alternative Catalog Number	HPS-IC-N-M-100-100
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Nitrite1000 ug/mL in H2O Exp Date: 1 Year, 12 months expiry date. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:2016 accredited....