



Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

**Telephone:** +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## **Product Datasheet**

## Nitrogen from Nitrate Anion IC Standard from Sodium Nitrate at 1,000 $\mu$ g/mL in H2O HPS-IC-NO3-M-100

Article Name	Nitrogen from Nitrate Anion IC Standard from Sodium Nitrate at 1,000 $\mu\text{g/mL}$ in H2O
Biozol Catalog Number	HPS-IC-NO3-M-100
Supplier Catalog Number	IC-NO3-M-100
Alternative Catalog Number	HPS-IC-NO3-M-100-100
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Nitrogen from NaNO3 1000 ug/mL in H2O 19 months expiry date. Traceable to NIST 3185. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:20166 accredited. Can be used as stock solution for Nitrate as Nitrogen, EPA Method 9056. For reporting Nit