

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### Nitrate Anion IC Standard from Sodium Nitrate at 1,000 µg/mL in H<sub>2</sub>O HPS-IC-NO-M-100

Article Name	Nitrate Anion IC Standard from Sodium Nitrate at 1,000 µg/mL in H <sub>2</sub> O
Biozol Catalog Number	HPS-IC-NO-M-100
Supplier Catalog Number	IC-NO-M-100
Alternative Catalog Number	HPS-IC-NO-M-100-100
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Nitrate1000 ug/mL in H <sub>2</sub> O 18 months expiry date. Traceable to NIST 3185. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:2016 accredited. Can be used as stock solution for nitrate as nitrate, EPA Method 9056. For reporting Nitrogen using a ...