

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

7-component Initial Verification Graphite Furnace Calibration Standard Contains: HPS-ICV-GFAA-100

| | |
|----------------------------|--|
| Article Name | 7-component Initial Verification Graphite Furnace Calibration Standard Contains: |
| Biozol Catalog Number | HPS-ICV-GFAA-100 |
| Supplier Catalog Number | ICV-GFAA-100 |
| Alternative Catalog Number | HPS-ICV-GFAA-100-100 |
| Manufacturer | High Purity Standards |
| Category | Biochemikalien |
| Product Description | Graphite Furn Cal Std in 2% HNO ₃ , (µg/mL) Arsenic, Lead, Selenium, Thallium at 2, Cadmium, Chromium, Silver at 0.4 in 2% HNO ₃ . 12 months expiry date. Traceable to NIST 31XX series. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:2016 accre... |