

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### Lead (Pb) ICP-MS Standard at 10 µg/mL in 2% HNO<sub>3</sub> HPS-10-28-1-250

Article Name	Lead (Pb) ICP-MS Standard at 10 µg/mL in 2% HNO <sub>3</sub>
Biozol Catalog Number	HPS-10-28-1-250
Supplier Catalog Number	10-28-1-250
Alternative Catalog Number	HPS-10-28-1-250-250
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Pb 10 µg/mL in 2% HNO <sub>3</sub> 12 months expiry date. Traceable to NIST SRM 3128. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:20166 accredited....