

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Chlorate Anion IC Standard from Sodium Chlorate at 100 µg/mL in H2O HPS-IC-CLO3-250

Article Name	Chlorate Anion IC Standard from Sodium Chlorate at 100 µg/mL in H2O
Biozol Catalog Number	HPS-IC-CLO3-250
Supplier Catalog Number	IC-CLO3-250
Alternative Catalog Number	HPS-IC-CLO3-250-250
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Chlorate 100 µg/mL in H2O Exp Date: 1 Year 18 months expiry date. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:2016 accredited....