

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### Nitrate Anion IC Standard from Sodium Nitrate at 100 µg/mL in H2O HPS-IC-NO-250

Article Name	Nitrate Anion IC Standard from Sodium Nitrate at 100 µg/mL in H2O
Biozol Catalog Number	HPS-IC-NO-250
Supplier Catalog Number	IC-NO-250
Alternative Catalog Number	HPS-IC-NO-250-250
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Nitrate100 ug/mL in H2O 18 months expiry date. Traceable to NIST 3185. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:20166 accredited....