

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### 15-component ICP Wavelength Calibration Standard Contains: HPS-WAVECAL-2-250

Article Name	15-component ICP Wavelength Calibration Standard Contains:
Biozol Catalog Number	HPS-WAVECAL-2-250
Supplier Catalog Number	WAVECAL-2-250
Alternative Catalog Number	HPS-WAVECAL-2-250-250
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	ICP wavelength calibration Std in 5% HNO <sub>3</sub> , (µg/mL) Aluminium, Arsenic, Barium, Cadmium, Cobalt, Chromium, Copper, Manganese, Molybdenum, Nickel, Lead, Selenium, Strontium, Zinc at 50, Potassium at 500 in 5% HNO <sub>3</sub> . 12 months expiry date. Traceable to N...