

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Nitrogen from Nitrite Anion IC Standard from Sodium Nitrite at 1,000 µg/mL in H₂O **HPS-IC-NO₂-M-50**

Article Name	Nitrogen from Nitrite Anion IC Standard from Sodium Nitrite at 1,000 µg/mL in H ₂ O
Biozol Catalog Number	HPS-IC-NO ₂ -M-50
Supplier Catalog Number	IC-NO ₂ -M-50
Alternative Catalog Number	HPS-IC-NO ₂ -M-50-50
Manufacturer	High Purity Standards
Category	Biochemikalien
Product Description	Nitrogen from NaNO ₂ 1000 ug/mL in H ₂ O 18 months expiry date. ISO 9001:2015 certified, ISO/IEC 17025:2017 and ISO 17034:2016 accredited. Can be used as stock solution for EPA Methods 353.2, SM-4500-NO ₂ B and 300.1. Can be used as stock solution for Ni...