

Diagnostica Vertrieb GmbH, Leipziger Straße 4 85386 Eching, Germany **Telephone:** +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99 **E-Mail:** info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## **Product Datasheet**

## AMPIPROBE HCV assay kit ENZ-ENZ-GEN200-0100

Article Name	AMPIPROBE HCV assay kit
Biozol Catalog Number	ENZ-ENZ-GEN200-0100
Supplier Catalog Number	ENZ-GEN200-0100
Alternative Catalog Number	ENZ-ENZ-GEN200-0100-100
Manufacturer	Enzo Life Sciences
Category	Kits/Assays
Species Reactivity	Virus
Product Description	AMPIPROBE HCV assay kit
Sensitivity	In a validation study approved by the New York State Department of Health, the AMPIPROBE HCV assay was determined to have the following sensitivity: LOD serum - 5.46 IU/mL via 95% Probit analysis LOD plasma - 7.91 IU/mL via 95% Probit analysis LOQ serum
Target	Hepatitis virus
Application Notes	The AMPIPROBE HCV assay is compatible with any properly calibrated qPCR thermal cycler capable of exciting and reading emissions from FAM/Green and CY5/Red. It has been validated for use on the following instruments: QIAGEN Rotor-Gene Q and Applied Bios