



Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

AMPIPROBE Candida 2 Assay Kit ENZ-ENZ-GEN203-0100

Article Name	AMPIDDORE Candida 2 Accay Kit
Article Name	AMPIPROBE Candida 2 Assay Kit
Biozol Catalog Number	ENZ-ENZ-GEN203-0100
Supplier Catalog Number	ENZ-GEN203-0100
Alternative Catalog Number	ENZ-ENZ-GEN203-0100-100
Manufacturer	Enzo Life Sciences
Category	Kits/Assays
Application	PCR
Product Description	AMPIPROBE Candida 2 Assay Kit
Sensitivity	In a validation study approved by the New York State Department of Health, the AMPIPROBE Candida 2 Assay was determined to have the following percentage of Sensitivity: Candida parapsilosis - 94.70% Candida tropicalis - 100.10%
Target	Candida parapsilosis Candida tropicalis
Application Notes	The AMPIPROBE Candida 2 Assay is compatible with any properly calibrated qPCR thermal cycler capable to detect fluorescence decay. It has been validated for use on the QIAGEN Rotor-Gene Q