



Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

AMPIPROBE Bacterial Vaginosis Assay Kit ENZ-ENZ-GEN205-0100

Article Name	AMPIPROBE Bacterial Vaginosis Assay Kit
Biozol Catalog Number	ENZ-ENZ-GEN205-0100
Supplier Catalog Number	ENZ-GEN205-0100
Alternative Catalog Number	ENZ-ENZ-GEN205-0100-100
Manufacturer	Enzo Life Sciences
Category	Kits/Assays
Species Reactivity	Bacteria
Product Description	AMPIPROBE Bacterial Vaginosis Assay Kit
Sensitivity	In a validation study approved by the New York State Department of Health, the AMPIPROBE Bacterial Vaginosis Assay was determined to have a sensitivity of 84.30%.
Target	Atopobium vaginae Gardnerella vaginalis Lactobacillus spp. Megasphaera-1 BVAB-2
Application Notes	The AMPIPROBE Bacterial Vaginosis Assay is compatible with any properly calibrated qPCR thermal cycler capable to detect fluorescence decay. It has been validated for use on the QIAGEN Rotor-Gene Q