

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

AMPIPROBE NG/CT/TV Assay Kit ENZ-ENZ-GEN207-0100

Article Name	AMPIPROBE NG/CT/TV Assay Kit
Biozol Catalog Number	ENZ-ENZ-GEN207-0100
Supplier Catalog Number	ENZ-GEN207-0100
Alternative Catalog Number	ENZ-ENZ-GEN207-0100-100
Manufacturer	Enzo Life Sciences
Category	Kits/Assays
Species Reactivity	Bacteria, Parasite
Product Description	AMPIPROBE NG/CT/TV Assay Kit...
Sensitivity	In a validation study approved by the New York State Department of Health, the AMPIPROBE NG/CT/TV Assay was determined to have the following percentage of Sensitivity: Neisseria gonorrhoeae - 100.00% Chlamydia trachomatis - 100.00% Trichomonas vaginalis
Target	Neisseria gonorrhoeae Chlamydia trachomatis Trichomonas vaginalis
Application Notes	The AMPIPROBE NG/CT/TV Assay is compatible with any properly calibrated qPCR thermal cycler capable to detect fluorescence decay. It has been validated for use on the QIAGEN Rotor-Gene Q