

Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Magnetic Plus DNA/RNA Auto Extraction kit (Form P1) - CE IVD ATB-80-09-P1

| Article Name | Magnetic Plus DNA/RNA Auto Extraction kit (Form P1) - CE IVD |
|----------------------------|--|
| Biozol Catalog Number | ATB-80-09-P1 |
| Supplier Catalog Number | 80-09-P1 |
| Alternative Catalog Number | ATB-80-09-P1 |
| Manufacturer | Astra Biotech GmbH |
| Category | Kits/Assays |
| Species Reactivity | Human |
| Product Description | simultaneous isolation of viral, bacterial, and human DNA and RNA from swabs (buccal, nasopharyngeal and oropharyngeal), clinical samples of plasma, whole blood, dried blood spots, urine, leukocyte fraction, mucous membranes of the urogenital tract (|
| Samples | Swabs (buccal, nasopharyngeal and oropharyngeal), clinical samples of plasma, whole blood, dried blood spots, urine, leukocyte fraction, mucous membranes of the urogenital tract (urethra, cervical canal and vagina), saliva, bronchial lavage (bronchoalveo |
| Application Notes | Guanidine thiocyanate cell lysis and subsequent sorption of nucleic acids on surface of magnetic beads, followed by washing steps and elution of purified nucleic acids. Pre-filled into deep-well-plates. Automated. Validated on KingFisher(TM) Flex (Thermo Fi |