



Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

**Telephone:** +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## **Product Datasheet**

## CLIA Kit for Luteinizing Hormone (LH) USC-CCA441RA

Article Name	CLIA Kit for Luteinizing Hormone (LH)
Biozol Catalog Number	USC-CCA441RA
Supplier Catalog Number	CCA441Ra
Alternative Catalog Number	USC-CCA441RA-48, USC-CCA441RA-96
Manufacturer	Cloud-Clone
Category	Kits/Assays
Species Reactivity	Rat
Range	31.25-8,000pg/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 11.73pg/mL
Samples	serum, plasma and other biological fluids
Application Notes	Chemiluminescent immunoassay for Antigen Detection.