



Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

CLIA Kit for Aprotinin (AP) USC-CCA968BO

Article Name	CLIA Kit for Aprotinin (AP)
Biozol Catalog Number	USC-CCA968BO
Supplier Catalog Number	CCA968Bo
Alternative Catalog Number	USC-CCA968BO-48, USC-CCA968BO-96
Manufacturer	Cloud-Clone
Category	Kits/Assays
Species Reactivity	Bovine
Range	1.95-500ng/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 0.71ng/mL
UniProt	P00974
Samples	serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Application Notes	Chemiluminescent immunoassay for Antigen Detection.