



Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

ELISA Kit for Thyroxine (T4) USC-CEA452GE

Article Name	ELISA Kit for Thyroxine (T4)
Biozol Catalog Number	USC-CEA452GE
Supplier Catalog Number	CEA452Ge
Alternative Catalog Number	USC-CEA452GE-48, USC-CEA452GE-96
Manufacturer	Cloud-Clone
Category	Kits/Assays
Species Reactivity	AII
Range	3.70-300ng/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 1.29ng/mL
Samples	serum, plasma and other biological fluids
Application Notes	Enzyme-linked immunosorbent assay for Antigen Detection.