

Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

**Telephone:** +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## **Product Datasheet**

## High Sensitive ELISA Kit for Follicle Stimulating Hormone (FSH) USC-HEA830GA

| Article Name               | High Sensitive ELISA Kit for Follicle Stimulating Hormone (FSH)         |
|----------------------------|---|
| Biozol Catalog Number      | USC-HEA830GA  |
| Supplier Catalog Number    | HEA830Ga  |
| Alternative Catalog Number | USC-HEA830GA-48, USC-HEA830GA-96  |
| Manufacturer               | Cloud-Clone   |
| Category                   | Kits/Assays   |
| Species Reactivity         | Gallus  |
| Range                      | 15.6-1,000pg/mL   |
| Sensitivity                | The minimum detectable dose of this kit is typically less than 6.3pg/mL |
| Samples                    | Serum, plasma and other biological fluids                               |
| Application Notes          | Enzyme-linked immunosorbent assay for Antigen Detection.                |