



Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

**Telephone:** +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## **Product Datasheet**

## CLIA Kit for Meningioma Expressed Antigen 5 (MGEA5) USC-SCA069MU

Article Name	CLIA Kit for Meningioma Expressed Antigen 5 (MGEA5)
Biozol Catalog Number	USC-SCA069MU
Supplier Catalog Number	SCA069Mu
Alternative Catalog Number	USC-SCA069MU-48, USC-SCA069MU-96
Manufacturer	Cloud-Clone
Category	Kits/Assays
Species Reactivity	Mouse
Range	68.6-50,000pg/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 29.4pg/mL
UniProt	Q9EQQ9
Samples	Urine
Application Notes	Chemiluminescent immunoassay for Antigen Detection.