



Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

CLIA Kit for Matrix Metalloproteinase 13 (MMP13) USC-SCA099RB

Article Name	CLIA Kit for Matrix Metalloproteinase 13 (MMP13)
Biozol Catalog Number	USC-SCA099RB
Supplier Catalog Number	SCA099Rb
Alternative Catalog Number	USC-SCA099RB-48, USC-SCA099RB-96
Manufacturer	Cloud-Clone
Category	Kits/Assays
Species Reactivity	Rabbit
Range	0.34-250ng/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 0.15ng/mL
UniProt	O62806
Samples	Serum, plasma and other biological fluids
Application Notes	Chemiluminescent immunoassay for Antigen Detection.