



Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

CLIA Kit for Matrix Metalloproteinase 8 (MMP8) USC-SCA103HU

| Article Name | CLIA Kit for Matrix Metalloproteinase 8 (MMP8) |
|----------------------------|--|
| Biozol Catalog Number | USC-SCA103HU |
| Supplier Catalog Number | SCA103Hu |
| Alternative Catalog Number | USC-SCA103HU-48, USC-SCA103HU-96 |
| Manufacturer | Cloud-Clone |
| Category | Kits/Assays |
| Species Reactivity | Human |
| Range | 3.43-2,500pg/mL |
| Sensitivity | The minimum detectable dose of this kit is typically less than 1.34pg/mL |
| UniProt | P22894 |
| Samples | Serum, plasma, tissue homogenates, cell culture supernates and other biological fluids |
| Application Notes | Chemiluminescent immunoassay for Antigen Detection. |