

Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

CLIA Kit for Xeroderma Pigmentosum, Complementation Group C (XPC) USC-SCA473HU

Article Name	CLIA Kit for Xeroderma Pigmentosum, Complementation Group C (XPC)
Biozol Catalog Number	USC-SCA473HU
Supplier Catalog Number	SCA473Hu
Alternative Catalog Number	USC-SCA473HU-48, USC-SCA473HU-96
Manufacturer	Cloud-Clone
Category	Kits/Assays
Species Reactivity	Human
Range	13.7-10,000pg/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 5.4pg/mL
UniProt	Q01831
Samples	Tissue homogenates and other biological fluids
Application Notes	Chemiluminescent immunoassay for Antigen Detection.