



Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

CLIA Kit for Xanthine Dehydrogenase (XDH) USC-SCC608BO

Article Name	CLIA Kit for Xanthine Dehydrogenase (XDH)
Biozol Catalog Number	USC-SCC608BO
Supplier Catalog Number	SCC608Bo
Alternative Catalog Number	USC-SCC608BO-48, USC-SCC608BO-96
Manufacturer	Cloud-Clone
Category	Kits/Assays
Species Reactivity	Bovine
Range	1.37-1,000pg/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 0.58pg/mL
UniProt	P80457
Samples	Serum, plasma and other biological fluids
Application Notes	Chemiluminescent immunoassay for Antigen Detection.