

Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

**Telephone:** +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## **Product Datasheet**

## Human Exostoses Like Protein 1 (EXTL1) CLIA Kit ABX495930

Article Name	Human Exostoses Like Protein 1 (EXTL1) CLIA Kit
Biozol Catalog Number	ABX495930
Supplier Catalog Number	abx495930
Alternative Catalog Number	ABX495930-96TESTS, ABX495930-5X96TESTS, ABX495930- 10X96TESTS
Manufacturer	Abbexa
Category	Kits/Assays
Application	CLIA
Species Reactivity	Human
Range	0.312 ng/ml - 20 ng/ml
Sensitivity	< 0.115 ng/ml
Samples	Tissue homogenates and other biological fluids.
Target	Exostoses Like Protein 1 (EXTL1)