

Diagnostica Vertrieb GmbH, Leipziger Straße 4 85386 Eching, Germany **Telephone:** +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99 **E-Mail:** info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## **Product Datasheet**

## Progesterone ELISA BVD-RCD020R

Article Name	Progesterone ELISA
Biozol Catalog Number	BVD-RCD020R
Supplier Catalog Number	RCD020R
Alternative Catalog Number	BVD-RCD020R-96
Manufacturer	BioVendor
Category	Kits/Assays
Application	Serum
Species Reactivity	Human
Product Description	Competitive ELISA, Immobilized antigen
Range	0.3-60 ng/ml
Sensitivity	0.1 ng/ml
Tag	BioVendor
Storage	Store the complete kit at 2–8°C. Under these conditions, the kit is stable until the expiration date (see label on the box).
Target	Progesterone
Application Notes	Intra assay: $CV = 10.4\%$ . Inter assay: $CV = 11.4\%$ . Spiking recovery: 98,00%. Dilution linearity: 93,00%