



Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

**Telephone:** +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## **Product Datasheet**

## **QuantiFluo(TM) Diamine Oxidase Assay Kit BAS-QFDO-100**

Article Name	QuantiFluo(TM) Diamine Oxidase Assay Kit
Biozol Catalog Number	BAS-QFDO-100
Supplier Catalog Number	QFDO-100
Alternative Catalog Number	BAS-QFDO-100
Manufacturer	BioAssay Systems
Category	Kits/Assays
Product Description	For quantitative determination of diamine oxidase and evaluation of drug modulators. Key Features: Fast and sensitive. Use of 10 $\mu$ L sample. Linear detection range 0.5 to 6 U/L for 30 min reaction at 25C. Convenient. The procedure involves adding a si