



Diagnostica Vertrieb GmbH, Leipziger Straße 4

85386 Eching, Germany

Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99

E-Mail: info@biozol.de

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Pyruvate Assay Kit CBL-MET-5029

Article Name	Pyruvate Assay Kit
Biozol Catalog Number	CBL-MET-5029
Supplier Catalog Number	MET-5029
Alternative Catalog Number	CBL-MET-5029
Manufacturer	Cell Biolabs
Category	Kits/Assays
Product Description	Our Pyruvate Assay Kit measures pyruvate in biological samples. First, pyruvate is oxidized by pyruvate oxidase, producing hydrogen peroxide. The hydrogen peroxide is then detected at ex. 530-570 nm/em. 590-600 nm using a specific fluorometric probe